

ANALYSIS CERTIFICATE

Product Name: American Cardioan Gel 250ml

Analysis Date: 15.11.25

Lot Number: ARU-11251

Production Date: 11.11.2025

Expiry Date: 30.10.2028

ANALYSIS PARAMETERS

Physical and Chemical Analysis

Colour

Colorless

-

Smell

Odorless

-

Appearance

Gel

-

Density (23° C)

0.925 gr/cm³

ASTM D 891-89

Viscosity (23° C, 6sp, 5rpm)

164250 Cp

Acmas Viscometer

Ph. (23° C)

6.19

pH Meter

RESULT:

The sample referred to above is of standard quality as defined in the mentioned references made there under.

----- End of Report-----

Date of Completion: 17.11.2025

Analyzed By:



Note:

1. The results listed refer only the tested samples and applicable parameter. Endorsement of product is either inferred not implied.
2. This report is not to be reproduced wholly or in part; cannot be used as evidence in the court of law and should not be used in advertising media without special permission in writing.