

ANALYSIS CERTIFICATE

Product Name: **CARDIOCAN ELECTRODE\ ULTRASOUND "250.ML" and JAR "5LTR"**

Analysis Date: 17.06.2024

BATCH NO: UE-006

Production Date: 18.06.2024

Expiry Date: 05/27.

ANALYSIS PARAMETERS RESULTS REFERENCES Physical and Chemical Analysis.

| | |
|-------------------------------|---|
| Colour | : Blue |
| Smell | : Odorless |
| Appearance | : Gel |
| Density | :(23° C) 0,981 gr/cm ³ ASTM D 891-89 |
| Viscosity :(23° C, 6sp, 5rpm) | :164000 Cp Brookfield DV-II+Viscometer |
| Ph. | :(23° C) 6,61 Inolab pH Meter Microbiology |

Total Aerobic Mesophilic Bacteria < 10 cfu/ml- Escherichia Coli 0 cfu/ml-g Staphylococcus Aureus 0 cfu/ml- Pseudomonas Aeruginosa 0 cfu/ml-g Total Yeast and Mold < 10 cfu/ml- Candida Albicans 0 cfu/ml-.

RESULT: The analysis has been completed on the "CARICON ELECTRODE/ ULTRASOUND GEL "250ml" AND "5Ltr." a sample which was delivered 17.06.2024 the results are approved.

MAX INDUSTRIES.
44,45 UDYOG NAGAR
PALDA(INDORE)
INDORE MP .452016



