50, Gajanan Industrial Park, Nemawar Road, Gram Asrawad Bujurg,
Indore (M.P.) – 452016

MEDWIN PHARMATECH
A healthy you

ANALYSIS CERTIFICATE

Product Name: American Cardiocan Gel 5000ml
Analysis Date: 25.09.26
Lot Number: ARU-09261
Production Date: 23.09.2026
Expiry Date: 31.08.2029

ANALYSIS PARAMETERS	RESULTS	REFERENCES
Physical and Chemical Analysis		
Colour	Colorless	-
Smell	Odorless	-
Appearance	Gel	-
Density (23º C)	0.943 gr/cm^3	ASTM D 891-89
Viscosity (23º C, 6sp, 5rpm)	168243 Cp	Acmas Viscometer
Ph. (23º C)	6.87	pH Meter

RESULT:

The sample referred to above is of standard quality as defined in the mentioned references made there under.

------- End of Report-------

Date of Completion: 26.09.2026

Analyzed By:

Note:
1. The results listed refer only the tested samples and applicable parameter. Endorsement of product is either inferred not implied.
2. This report is not to be reproduced wholly or in part; cannot be used as evidence in the court of law and should not be used in advertising media without special permission in writing.